



Operators Manual

Symbols on the unit and case

	Caution! (electrical output).
	Follow operating instructions! Failure to do so could place the patient or operator at risk.
	TENS should not be used by those fitted with a demand style pacemaker, unless directed to do so by a healthcare provider.
	Patient's shock protection type: BF (Body Floated) Equipment. This equipment is not earthed but contains a battery within an insulated unit.
	Indicates the manufacturer's catalogue number so that the medical device can be identified.
	Manufacturer's LOT/Batch number. Present it together with SN number when you report a technical fault or claim a warranty return.
	Manufacturer's serial number of the unit. Present it together with LOT number when you report a technical fault or claim a warranty return.
	Name and address of Manufacturer.
	Date of manufacture.
	Conformity indication with the essential health and safety requirements set out in European Directives. 0088 - Notified Body identification (LRQA Ltd.)
	This product should be kept dry.
IP20 on the unit	This is an indication for protection against ingress of water and particulate matter. The mark IP20 on your unit means: your unit is protected against solid foreign objects of 12.5mm dia and greater. Not protected against water.
IP02 on the case	IP02 on the carrying case means: Protected from the ingress of water droplets from a shower of rain.
	Do not dispose in normal dustbin (see page 19 for the disposal instructions).

Warnings

- * This unit must be used with the guidance of a Physiotherapist or Doctor.
- * Type BF equipment, Continuous Operation.
- * Do not insert lead wires into a mains power supply.
- * Do not immerse unit into water or any other substance.
- * The unit is not protected from the ingress of water droplets from a shower or rain if used outside the carrying case.
- * Do not use the TENS unit in the presence of a flammable anaesthetic gas mixture and air or with Oxygen or Nitrous Oxide.
- * If using rechargeable 9 Volt PP3 Nickel Metal Hydride batteries, be sure to use a CE approved battery charger. Never connect the Universal TENS directly to a battery charger or to any other mains powered equipment.
We advise not to use Ni-Cad rechargeable batteries.
- * Skin electrode pads are for single patient use only.
- * Keep out of reach of children.
- * Do not use this TENS machine on your face unless you are under strict guidance from a qualified Clinician.
- * Application of electrodes near the thorax may increase the risk of cardiac fibrillation.
- * Operation in close proximity (e.g. 1m) to a shortwave or microwave therapy equipment may produce instability in the stimulator output.
- * Simultaneous connection of a patient to high frequency surgical equipment may result in burns at the site of the stimulator electrodes and possible damage to the stimulator.
- * No modification of this equipment is allowed!

Contents

Contents:	Page:
Unit symbols explained	2
Warnings	3
Intended Purpose	5
What is Pain?	5
What is TENS?	5
Contra Indications & Precautions	6
Description of Universal TENS Unit & Functions	7
Quick Start Instructions	8
Programmes	9
Lock Mode Function	10
Using the Universal TENS Unit	11
Treatment Modes	12
How Long do I Use Universal TENS For?	12
Electrode Placement	13
Dermatomes & Myotomes	13
Contiguous Placement	13
Acupuncture Points	13
Electrode Types & Tips	14
Suggested Electrode Placement	15
Care, Maintenance, Accessories and Disposal	19
Indications for use	20
Specifications	21
Information regarding Electromagnetic compatibility and interference (EMC)	22
Troubleshooting	26
Commonly Asked Questions	27
Warranty	28
Dermatome Charts	29
Clinical References	31

Intended Purpose

Universal TENS uses a small battery operated unit to provide a non-invasive, drug free method of controlling acute and principally long term intractable pain. It can also be used as adjunctive treatment in the management of post surgical traumatic pain problems. Mild electrical impulses are transmitted through the skin via surface electrodes to modify the body's pain perception.

What is Pain?

When we feel pain it is the body's process of informing us that something is wrong. To feel pain is important, without this feeling abnormal conditions may go undetected, creating damage or injury to critical parts of the body.

Although pain is essential in warning our body of trauma or malfunction, nature may have gone too far in its design. Continued long-term chronic pain has no useful value apart from its importance in diagnosis. Pain begins when a coded signal travels to the brain where it is decoded, and analysed. The pain message travels from the injured area of the body along small diameter nerves leading to the spinal cord. At this point the message is switched to a different kind of nerve that travels up the spinal cord to the brain area. The brain then analyses the pain message, refers it back and the pain is felt.

What is TENS?

Transcutaneous Electrical Nerve Stimulation (TENS) uses a small battery operated unit to provide a non-invasive, drug free method of controlling acute and principally long term intractable pain. It can also be used as an adjunctive treatment in the management of post surgical traumatic pain problems. In TENS mild electrical impulses are transmitted through the skin via surface electrodes to modify the body's pain perception. Universal TENS does not cure problematic physiological conditions; it only helps to control the pain perception. Universal TENS will not work for every user. Please seek advice from your Doctor.

There are millions of small nerve fibres throughout the body and it only requires a few impulses to produce chronic pain. In addition to small fibres, which allow the sensation of pain to be felt, the body is also made up of larger diameter nerve fibres. These larger nerve fibres transmit less unpleasant sensations such as touch or warmth, assisting us to form an impression of our environment. Stimulating the larger nerve fibres using Universal TENS may have the effect of inhibiting the transmission of pain along the smaller nerve fibres to the spinal cord [known as the 'Pain Gate Theory'].

Contra Indications & Precautions


Before using this equipment you must first seek the advice of your Physiotherapist or Doctor.

Please read this manual before using the Universal TENS machine.



Universal TENS should not be used:

- * By patients fitted with a demand style cardiac pacemakers unless so advised by their Doctor.
- * During pregnancy [unless medically advised].
- * By patients with undiagnosed pain conditions.
- * By patients with undiagnosed skin conditions.
- * With patients who have diminished mental capacity or physical competence who cannot handle the device properly.
- * On anaesthetised or desensitised skin.
- * When driving a vehicle or operating potentially dangerous equipment.
- * Do not place electrodes:
 - > Over carotid sinus nerves.
 - > Over larynx or trachea.
 - > Inside the mouth.
 - > Over the area of the heart unless so advised by your Doctor.
 - > On your face unless under strict guidance from a qualified Clinician.
 - > Do not apply stimulation across or through the head, directly on the eyes, covering the mouth, on the front of the neck (especially the carotid sinus) or via electrodes placed on the chest and upper back or crossing over the heart.
- * The patient should use the unit only as prescribed.
- * Do not immerse the unit in water or any other liquid.
- * If you experience skin irritation this may be due to over-stimulation. In this case leave the skin to heal and use Universal TENS only for the periods prescribed. Turning the current up too high can cause skin irritation. In this case allow the skin to heal and use Universal TENS at a lower intensity. Some people experience an allergic reaction to the adhesive coating on the surface of the electrode. If this happens use a different make of electrode or change the electrode. If it continues try reducing the pulse width. If the problem still persists try moving the electrode position each day by just the width of the electrode, making sure the electrode positioning is still over the dermatome.
- * Keep unit out of reach of children.
- * Only use CE approved skin electrodes.
- * If in doubt about the use of the Universal TENS unit, call your Doctor, Therapist, Clinician or your distributor for advice.

Description of Universal TENS Unit & Functions



Concealed button locks the selected programme and records the patient's home compliance. Use a metal pin of the lead wire to press the concealed button located in the battery compartment.



- * **PRG button** Selects the desired set programme from P01 - P11 or customised programme PC1 - PC3.
- * **SET button** This button works only for programmes PC1-PC3. Press and hold SET button for 3 seconds to set parameters of your custom treatment: Pulse Rate, Pulse Width, Time, etc.
- * **ESC button** Stores customised programme and returns to the home position.

Quick Start Instructions

1. Insert a 9 volt PP3 Alkaline battery. Alternatively insert a rechargeable Nickel Hydride battery [Which is safer and has a much longer life than the Ni-Cad rechargeable batteries] into the battery compartment.
2. Insert lead wire/s to channel A and B if both channels are to be used.
3. Switch on the unit by pressing the ON/OFF button
4. Press the PRG [Programme] button to select one of the programmes as detailed in table 1 and table 2 on page 9.
5. To start press channel A + and B + button if you are using both channels.
6. To stop the programme, press the ON/OFF button which will turn the unit off.

Setting up your own continuous mode parameters for PC 1 or PC 2.

1. Select PC1 or PC2 by pressing the PRG button on the front panel. Press and hold the SET button for 3 seconds and the Hz symbol will flash, then press the + or – button on the front panel to adjust the Pulse Rate (frequency) from 2 - 200 Hz.
2. Press the SET button again and the μ S symbol will flash, then press the + or – button to adjust the Pulse Duration from 50 - 300 μ S.

Setting up your own Modulated mode parameters for PC3.

1. Select PC3 by pressing the PRG button on the front panel. Press and hold the SET button for 3 seconds until FHI is displayed. Press the + or - button on the front panel to adjust the high pulse rate (frequency) from 2 - 200 Hz.
2. Press the SET button and FLO will display. Press the + or - button on the front panel to adjust the low pulse rate (frequency) from 2 - 200 Hz.
3. Press the SET button and μ S symbol will flash and WLO will display. Press the + or - button on the front panel to adjust the high pulse width from 50 - 300 μ S.
4. Press the SET button and WLO will display. Press the + or - button on the front panel to adjust the low pulse width from 50 - 300 μ S.
5. Press the SET button again and the Clock symbol will flash ON/OFF, then press the + or – button to adjust the time. Channel A + or – button to hours and Channel B + or – button to change minutes.
6. After setting up the customised programme parameters, press the ESC button to store the information. Simply repeating the above procedure can reprogramme customised programmes.

Note: You must press the ESC button before locking the unit.

Programmes

Table 1

Programme	P1	P2	P3	P4	P5	P6	P7	P8
Mode	CON	CON	CON	CON	CON	CON	MP/MF	MF
Pulse rate [Hz]	90	90	80	100	10	2	100/65	10/90
Pulse duration [μ S]	200	175	200	200	200	200	200/100	200
Time	4hr	4hr	4hr	4hr	4hr	4hr	2hr	2hr

Table 2

Programme	P9	P10	P11	PC1	PC2	PC3
Mode	BST	A=con B=BST	MF	CON	CON	MF/MP
Pulse rate [Hz]	see below	see below	2 & 100	2 to 100	2 to 100	Low: 2 to 100 High: 2 to 100
Pulse duration [μ S]	200	175	200	50 to 300	50 to 300	Low: 50 to 300 High: 50 to 300
Time	2hr	2hr	35min	cust	cust	cust

Notes:

Programme 7: Pulse width modulating from 200 μ S to 100 μ S to 200 μ S, Pulse rate modulating from 100 Hz to 65 Hz to 100 Hz over 3 seconds.

Programme 9: Burst of 9 pulses of 200 μ S at 150 Hz repeated 2 times per second.

Programme 10: Channel A Constant 80 Hz and 200 μ S.
Channel B Burst of 9 pulses of 200 μ S at 150 Hz repeated 2 times per second.

Programme 11: 2 Hz for 3 seconds, 100 Hz for 3 seconds, repeating.

Programme PC3: Pulse width modulating from high value to low value to high value, Pulse rate modulating from high value to low value to high value over 3 seconds.



Lock Mode Function

Lock Mode Function

A "concealed" Lock button is included in the Universal TENS unit, which allows the clinician to accurately monitor the "Home Compliance" of the patient between appointments. The lock function allows the device to be locked in two ways:- One {L:T} to measure the time in use over one hour, and the average mA current used, leaving the parameters i.e. Constant, Burst, Modulation and the Rate and Pulse Width to be freely altered by the user or alternatively {L:PT} Locking the device to measure, time, mA current used and locking the parameters in place, which then cannot be changed or altered by the patient during use.


Locking the Unit

Remove the battery cover and, using the end of the lead wire, gently press on the concealed lock button as shown in the diagram on page 6 until you hear a double bleep. {L:T} Lock time and Current will appear on the LCD screen. If you want to lock the parameters as well press the +/- button until {L:PT} appears. Press the ESC button to lock parameters in place.



To Unlock the Unit

To unlock the unit and display the lock information, remove the battery cover, using the end of the 2mm dia pin press the concealed switch once and you will here a single bleep, this indicates the unit is now unlocked. The information for time in use and the average m A current used can be read on the front of the LCD display as seen on the diagram below. When you have noted the information press the ESC button to bring the unit back to the Home position.



Using the Universal TENS Unit

RATE [Hz or pulses per second]

The **RATE** to be selected depends primarily on the electrode placement on the patient's body. If one uses contiguous and dermatome (the electrodes alongside or over the area of pain) electrode placement, a higher rate of 80 Hz –100 Hz is desirable. The patient should experience steady continuous stimulation. It has been found that an optimal setting of 80 or 90 Hz with a pulse width of 200 μ S has good effect for most patients and is a good first choice for pain-gating. Patients using Trigger, motor or acupuncture points tend to respond to low rate stimulation 2 Hz-10 Hz and pulse width of 200 μ S. The desired effect is for the patient to feel individual pulses.

PULSE WIDTH [Duration]

The wider pulse widths will deliver stronger stimulation for any given intensity [mA] setting. By using a combination of intensity and pulse duration, it is felt that various pulse widths are capable of stimulating different groups of nerve fibres. The wider pulse duration is needed to recruit motor fibres, where as the narrow pulse duration is used more on the sensory fibres. The selection of which pulse duration to use is dependent upon the intended treatment protocol.

Stimulating the larger nerve fibres is thought to reduce the speed and the amount at which information is transmitted along the smaller nerve fibres. Also under certain circumstances the brain is thought to produce its own analgesic pain-killing substances, known as endorphins or endogenous opiates.

Intensity [mA]

Patients respond differently to the level of intensity, this is due to differences in individual patient's skin resistance, innervation and the type and condition of electrode being used.

A good formula for setting the intensity is to increase the current so that the patient feels slight muscle contraction, but not strong enough to move a joint, and then slightly reduce the intensity so that it feels comfortable. When using low rate TENS settings, individual twitches will occur. The higher rate TENS settings will increase muscle tension. It is not advised to increase the intensity to experience strong muscle contraction.

Treatment Modes

There are three treatment modes available on the Universal TENS unit:

- 1. Conventional TENS or normal.** This mode enables the user to select any rate between 2 Hz – 200 Hz, and a pulse width between 50 μ S-300 μ S. This is the most frequently used of the three modes. The most common selection is 80 Hz with a 200 μ S pulse width.
- 2. Burst Mode.** This mode is comparable to the low rate TENS technique except that each low rate pulse is substituted for by a short BURST of 9 pulses [200 μ S] at 150 Hz. It is a combination of conventional and low rate TENS. The burst mode is often referred to as acupuncture - like TENS.
- 3. Modulation TENS** this mode was designed to help prevent nerve accommodation that some patient's experience. It is achieved by continuously cycling the pulse width and rate.

How Long do I use TENS For?

This depends on the individual patient's condition, accuracy of electrode placement, stimulation and the characteristics selected, but typically the onset of pain relief starts after 20-30 minutes. Generally Universal TENS is used for longer periods of normally 1 hour 30 minutes per session. With some patients it can be much longer.

Electrode Placement

The placement of electrodes is one of the most important parameters in achieving effective pain relief when using Universal TENS. This is best left to your Physiotherapist or Doctor to advise as to which location is most appropriate. It may transpire that various positions need to be experimented with before the user finds the most effective positioning. The positioning may be via the contiguous, dermatome, myotome, motor, trigger or acupuncture points.

Dermatomes & Myotomes

These are areas of the body enervated by a single nerve root via the spinal cord. Each nerve root serves a known area of the skin. The dermatomes are named after the nerve root which serves it. For details of dermatome sites refer to diagrams on pages 29& 30.

Contiguous Placement

This form of electrode placement is the most common method used. It involves placing the red lead [the cathode proximal i.e. nearer the spine or center of the body] alongside the spine where the dermatome [on which your pain lies] enters and exists. The black lead [distal] is normally placed over or near to the pain site. Your Physiotherapist or Doctor may direct the current to cross through the pain area or using the 'bracket' system allow the current to flow on either side of the pain site through the nerve branches that supply the pain location.



Acupuncture Points

The placement of the red and black electrodes on the skin forms the electrical circuit for Universal TENS. It is the skin itself that creates the highest electrical resistance to stimulation. The Physiotherapist or Doctor may consider using acupuncture loci, which offer much lower resistance properties, as a more effective site for placing the electrodes. Accurately locating an acupuncture point can be difficult, please seek advice from your Doctor or Physiotherapist.

Electrodes Types and Tips

- * Self-Adhesive reusable long-term electrodes (if looked after) have a typical life span of 4/6 weeks. We recommend cleaning the skin with an alcohol-based wipe before placing the electrodes. Please only use Universal TENS electrodes as these have been chosen for safety, efficacy and quality in mind. After use, place the electrodes back onto the plastic film and follow the care instructions.


Skin Electrode Types Available:

SHAPE	CODE	DESCRIPTION
	UT 4040	40 x 40 mm, square [** max 53mA]
	UT 5050	"50 x50 mm, square (recommended for general use)"
	UT 9050	90 x 50 mm, rectangular
	UT 10050	100 x 50 mm, rectangular
	UT 30	30mm diameter, round [** max 46mA]
** IMPORTANT : Don't use UT 4040 at more than 53mA and UT 30 at more than 46 mA.		

A Few Good Tips [Self- Adhesive Electrodes]


- * If you find the electrodes will not stick due to oily skin, cleanse the skin with soap and water, then rinse and dry the area around the electrode site. If this does not work, try cleansing the skin with a swab impregnated with alcohol.
- * Clip away hairy skin using scissors; don't use a razor to remove the hairs!
- * The electrodes conductive material is water- based. If it becomes saturated (e.g. from perspiration), it will lose its adhesive qualities. After use leave the electrodes face up overnight to dry out (replace on plastic film in the morning).
At some point the electrodes will become dry. Moisten the adhesive surface with a few drops of water, and apply onto the plastic film overnight. This procedure will increase the electrode life by few more days.

Suggested Electrode Placement





Pain caused by Finger Arthritis



+ = Red
- = Black



Pain caused by Knee Arthritis




Neuralgia of Trigeninus




Cervical (2 Positions)



If you are using electrodes on your face, we recommend you contact your physiotherapist or clinician for guidance




+ = *Red*
- = *Black*



Cephalalgia Overorbital



Mandibular Syndrome



Herpes Zoster




Phantom Limb





Back Pain



+ = *Red*
- = *Black*



Menstrual Pain



Lumbar Pain (2 Positions)




Tooth Ache




+ = Red
- = Black



Sciatic Pain (2 Positions)



Epicondylitis



Shoulder Pain



Feet Pain



Ankle Pain

Care, Maintenance, Accessories and Disposal

WARNING! Only medically approved accessories should be used!

CONTROL UNIT

- * Wipe the surface after each use with a damp cloth or antiseptic wipe or baby wipe.
- * Do not use cleaning sprays or alcohol based cleaning solutions
- * Control unit disposal: please return to Savantini Limited, details at the back of this booklet.

ACCESSORIES

Battery:

- * To change the battery, open the battery door on the rear of the control unit by pressing down on the raised rib pattern just below the belt clip. Lift the battery out of the compartment. This is very easy and can be done by the user.
- * Check periodically for any discharge from the battery
- * Remove battery completely from unit if not in use for any extended period of time (typically one week)
- * Low battery indicator of 6.9 volts is shown on LCD display, when flashing change the battery for a new one
- * Preferably use a PP3 alkaline battery
- * Battery disposal: please return to the supplier from whom you have purchased it.

Lead Wires:

- * The lead wires should be handled carefully and never stretched, as this can cause the stimulation to function below normal standards or not at all
- * Examine lead wires before each treatment for loose connections or damage
- * Avoid stretching and twisting the lead wires
- * Store the lead wires carefully after each use
- * Lead wires Disposal: please return to the supplier from whom you have purchased them.

Self-Adhesive Electrodes:

- * Check the short connectors have not become separated from the electrodes
- * Replace electrodes onto plastic film after use. If they drop onto the floor debris will adhere to conductive gel rendering the electrodes ineffective

Electrode life can be considerably reduced by:

- * The type and condition of the skin
- * Deep seated moisturisers or make-up

For the Best Results:

- * Before each use cleanse the skin
- * After each use stick the pads on the shiny insert card and store in a cool and dry place, such as the fridge. (not freezer).

Indications for use

- * back pain
- * phantom limb pain
- * postoperative pain
- * pain associated with knee arthritis

Specifications

Universal TENS

1. Dual channel: individually isolated circuits.
2. Amplitude: 0 - 80 mA into 500 Ohm load; indication only. Actual mA will tend to be less than indicated due to electrode impedance: at 1000 Ohms load (Electrodes in poor condition) the maximum will be limited to 70 mA, at 1500 Ohms load the maximum will be limited to 65 mA.
3. Type: Constant Current, maximum output voltage 180 Volts +10 / -30 Volts
4. Waveform: Asymmetrical, rectangular bi-phasic with zero DC current.
5. Selectable pulse width: 50 μ S -300 μ S [10% accuracy].
6. Pulse Rate selection: in the continuous mode 2 Hz - 200 Hz [5% accuracy].
7. Mode: Continuous, Burst or Modulated.
8. Burst mode: Bursts of 9 pulses [200 μ S] at 150 Hz, repeating twice every second.
9. Modulation mode: 6-second cycle of concurrent width modulation and pulse repetition rate modulation. Width starting at 200 μ S and decreasing exponentially to 100 μ S in three seconds and then returning back to 200 μ S in the next three seconds. Rate starting at 100 Hz, decreasing exponentially to 65 Hz and then returning to 100 Hz.
10. Time duration of the treatment selectable: 1 minute to 12 hours.
11. Battery: PP3 Alkaline, 9V.
Expected average battery life [of standard 800 mAh, alkaline]: 34 h.
12. Low Battery Indicator: If the battery goes below 6.9 volts +/- 0.2 volts the battery symbol will flash on/off once every second.
13. If the battery voltage is below 6.6(+/-0.2) volts the unit will not turn on.
14. Open Electrode Detect: If an open circuit is detected at the output of channel A or B the output current will be reset at zero.

Expected service life:

5 years. Careful use and maintenance extends the life of the unit over the service life limit.

Environmental Conditions for use:

+5 to +40 degrees Centigrade. 15-93% Humidity.

Environmental conditions for storage & transport:

-25 to +70 degrees Centigrade. 15-93% Humidity.

Physical dimensions: 119.2 x 69 x 28.7 mm

Weight: 0.07 kg without battery, 0.1 kg with battery.

Information regarding Electromagnetic Compatibility and Interference (EMC)

Universal TENS products are designed to produce very low levels of radio frequency (RF) emissions (interference), to be immune from effects of interference produced by other equipment operating in their vicinity and damage due to electrostatic discharge all when operating in a typical domestic and or clinical environment. They are certified to meet the international EMC standard EN60601-1-2. For more information please refer to the tables 201, 202, 204 and 206.

Table 201: Guidance and manufacturer's declaration
- electromagnetic emission

The Universal TENS unit is intended for use in the electromagnetic environment specified below. The customer or the user of the Universal TENS unit should ensure that it is used in such environment

Emission test	Compliance	Electromagnetic environment guidance
RF emission CISPR 11	Group 1	Universal TENS uses RF energy only for its internal function. Therefore, its RF emissions are very low and are not likely to cause any interference in nearby electronic equipment
RF emission CISPR 11	Class B	Universal TENS is suitable for use in all establishments , including domestic establishments and those directly connected to the public low voltage power supply network that supplies buildings used for domestic purposes
Harmonic emissions IEC 61000-3-2	Not applicable	
Voltage fluctuations/ flicker emissions IEC 61000-3-3	Not applicable	

Table 202: Guidance and manufacturer's declaration
– electromagnetic immunity

Universal TENS is intended for use in the electromagnetic environment specified below. The customer or the user of the Brand® product should assure that it is used in such an environment, and that precautions regarding that environment are heeded.

Immunity test	IEC 60601 test level	Compliance level	Electromagnetic environment guidance
Electrostatic discharge (ESD) IEC 61000-4-2	6 kV contact ±8 kV air	±6 kV contact ±8 kV air	Floors should be wood, concrete or ceramic tile. If floors are covered with synthetic material, the relative humidity should be at least 30 %.
Power frequency (50/60 Hz) magnetic field IEC 61000-4-8	3 A/m	3 A/m	Power frequency magnetic fields should be at levels characteristic of a typical location in a typical commercial or hospital environment.

Table 204: Guidance and manufacturer's declaration
– electromagnetic immunity

Universal TENS is intended for use in the electromagnetic environment specified below. The customer or the user of the Universal TENS unit should assure that it is used in such an environment.

Immunity test	IEC 60601 test level	Compliance level	Electromagnetic environment guidance
Conducted RF IEC 61000-4-6	3 Vrms 150 kHz to 80 MHz	3 Vrms 150 kHz to 80 MHz	Portable and mobile RF communications equipment should be used no closer to any part of the Universal TENS, including cables, than the recommended separation distance calculated from the equation applicable to the frequency of the transmitter. Recommended separation distance $d = 1.2 \sqrt{P}$ 150 kHz to 80 MHz, $d = 1.2 \sqrt{P}$ 80 MHz to 800 MHz $d = 2.3 \sqrt{P}$ 800 MHz to 2.5GHz where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer and d is the recommended separation distance in meters (m). Field strengths from fixed RF transmitters, as determined by an electromagnetic site survey ^a , should be less than the compliance level in each frequency range ^b . Interference may occur in the vicinity of equipment marked with the following symbol: 
Radiated RF IEC 61000-4-3	3 V/m 80 MHz to 2,5 GHz	3 V/m 80 MHz to 2,5 GHz	

NOTE 1 At 80 MHz and 800 MHz, the higher frequency range applies.

NOTE 2 These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.

^a Field strengths from fixed transmitters, such as base stations for radio (cellular/cordless) telephones and land mobile radios, amateur radio, AM and FM radio broadcast and TV broadcast cannot be predicted theoretically with accuracy. To assess the electromagnetic environment due to fixed RF transmitters, an electromagnetic site survey should be considered. If the measured field strength in the location in which the Universal TENS is used exceeds the applicable RF compliance level above, the Universal TENS unit should be observed to verify normal operation. If abnormal performance is observed, additional measures may be necessary, such as reorienting or relocating the Brand® product.

^b Over the frequency range 150 kHz to 80 MHz, field strengths should be less than 3 V/m.

Table 206: Recommended separation distances between portable and mobile RF communications equipment and the Universal TENS unit.

The Universal TENS unit is intended for use in an electromagnetic environment in which radiated RF disturbances are controlled. The customer or the user of the Universal TENS can help prevent electromagnetic interference by maintaining a minimum distance between portable and mobile RF communications equipment (transmitters) and the Universal TENS as recommended below, according to the maximum output power of the communications equipment.

Rated maximum output power of transmitter W	Separation distance according to frequency of transmitter		
	150 kHz to 80 MHz $d = 1.2 \sqrt{P}$	80 MHz to 800 MHz $d = \sqrt{1.2 P}$	800 MHz to 2,5 GHz $d = \sqrt{2.3 P}$
0,01	0.12	0.12	0.23
0,1	0.38	0.38	0.73
1	1.2	1.2	2.3
10	3.8	3.8	7.3
100	12	12	23

For transmitters rated at a maximum output power not listed above, the recommended separation distance d in meters (m) can be estimated using the equation applicable to the frequency of the transmitter, where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer.

NOTE 1 At 80 MHz and 800 MHz, the separation distance for the higher frequency range applies.

NOTE 2 These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.

Troubleshooting

Problem:

- Cannot reach maximum mA level; or
- The unit cuts off stimulation at a certain level; or
- When increasing the intensity 0 (zero) mA is flashing; or
- The power is cutting off during use.

Solution:

It is normal behaviour in our and any other quality muscle stimulators (and Universal TENS machines), and in most cases resolves itself - please read the guidance below.

The stimulation intensity will drop to zero if you simply press the mA+ button and no electrodes are connected to the channel on which you increase the intensity. You should attach a pair of electrodes to the lead wire and the lead should be connected to the channel on which you increase the stimulation in Universal TENSity (mA).

Our unit is designed to detect any poor or intermittent connection across the electrodes and to cut off the stimulation output (mA) when it does so. This is a safety precaution. It is designed to prevent the user from inadvertently turning up the output stimulation current in the presence of a poor or intermittent connection and then experiencing a large unexpected powerful surge in the stimulation, if and when the connection is re-established.

Reasons for no connection if you use surface skin electrodes:

- * Check if both electrodes are connected to the same dual conductor lead wire, one electrode to the black connector (-) and another to red connector (+).
- * Check if both electrodes are making a sticky contact on your skin, some electrode edges could not be stuck due to electrode wear & tear, but the electrode should be sticking with at least 80% of it's field. You may have dry lots of grease after long term use, try new electrodes. You may have dry gel on electrodes, try to make it more sticky by dropping a small amount of water on the black (conductive) side of the electrode and leave for an hour for the gel to absorb. Don't use wet electrodes! Try some fresh electrodes as electrodes loose conductivity proportionally to the use time due to grease and gel getting drier.
- * This is the most frequent reason: check if the dual conductor leadwire cable is not broken, as it might be bent or pulled out too much which results in no conductivity: try another cable. To check if the cable is good, cross the red and black pin and increase mA on the unit. If the cable conducts the electricity, the mA will go above 10mA and you would feel the stimulation mild tickling in your fingers which holds the crossed pins. If you feel a mild electrical current, this means the problem is with surface skin electrodes.

Commonly Asked Questions

- Q - *Does Universal TENS work for all pain conditions and on all patients?***
A - There is significant variation between patients with similar pain conditions. However, it is known that Universal TENS does work in up to 70% of cases.
- Q - *How can I have a better chance of success?***
A - Seeking professional advice from your Physiotherapist or Doctor on how to best apply Universal TENS is the best answer we can give to this question.
- Q - *Are there circumstances in which Universal TENS should not be used?***
A - Yes. For undiagnosed pain; When using a cardiac pace maker; During pregnancy and other instances as fully detailed in this manual on page 6.
- Q - *How long will I have to use the Universal TENS stimulator?***
A - Some long term chronic pain sufferers may have to use a stimulator for extended periods of time, even years. Other conditions may only need a short period of treatment lasting weeks.
- Q - *If I have any medical or product queries how can I get help?***
A - Any clinical advice on the Universal TENS stimulator should be provided by your Physiotherapist or Doctor.

Warranty

The warranty to the original purchaser is that this product will be free from defects in the material, components and workmanship, for a period of 2 years from the date of purchase by the user.

If you have any query regarding this unit please call our helpline in the first instance. If Savantini Limited is satisfied that the product is defective return it to us and we will liaise with the UK manufacturer Verity Medical Limited. The liability of Verity Medical Ltd., under this limited product warranty does not extend to any misuse or abuse such as dropping or immersing the unit in water or other liquid substance or tampering with the unit or normal wear and tear. Any evidence of tampering will nullify this warranty.

Customer Service and Helpline:

Please contact your distributor for any customer service enquiries, including the warranty returns.

For assistance, if needed, in setting up, using or maintaining the unit, or report unexpected operation or events, or to purchase spare TENS skin electrodes, or TENS garments please telephone our helpline or contact your on-line Universal TENS stockist.

Distributed by: Savantini Ltd.
Savantini House, Foster Street
Stoneferry Road, Hull
HU8 8BT, UK
Customer Services Helpline
+44(0) 1482 888785

Manufacturer: Verity Medical Ltd., Unit 7, Upper Slackstead Farm, Farley Lane, Braishfield, Romsey, Hampshire SO51 0QL, United Kingdom.


This product is manufactured by Verity Medical Ltd.,
in compliance with the European Union Medical Device Directive
MDD93/42/EEC under the supervision of LRQA Ltd.,
(Lloyd's Register Quality Assurance Ltd),
Notified Body number 0088.

CE 0088


Verity Medical Ltd., is certified by LRQA Ltd., to the following
Quality Standards: ISO13485:2016

Dermatome Charts

Anterior View



Posterior View.



Clinical References

Please contact us for any clinical referencies of Universal TENS:
Helpline +44(0) 1482 888785

Not for sale or use in the USA

Distributor:

Savantini Ltd.

Savantini House, Foster Street

Stoneferry Road, Hull

HU8 8BT, UK

England

CUSTOMER SERVICE HELPLINE

+44 (0)1482 888785

*Call us! We're here to help you
Monday-Friday, 8am - 5pm.*

**Proudly made in
the United Kingdom**

Document revision info.:

LOT OPH300A-OM-EN01-23-01-2019

Universal TENS
manual (English)

